

# Disposable Dental Kit

## 一、 Product name, model and specification

Product name: Disposable dental kit

Model specification: Examination type, nursing type

## 二、 Name, address, production address, contact information and after-sales service unit of the registered person/manufacturer

1、 Name: Hunan Zhenheyikang Medical Instrument Co.,LTD

2、 Address: No.6 Building Jingxiang Energy, No.55 Xiaguang east road, Gaoxin district Xiangtan, Hunan

3、 Production address: NO.6 Building, NO.8 building, 3<sup>rd</sup> and 4<sup>th</sup> floor Jingxiang Energy, No.55 Xiaguang east road, Gaoxin district Xiangtan, Hunan, NO 19 Chutian Road, Yuetang District, Xiangtan City, Hunan

4、 Contact number: +86 17711627084

5、 After-sales service unit: Hunan Zhenheyikang Medical Instrument Co.,LTD

三、 **Manufacturing License NO:** Hunan Medical Products Administration Production Licence No. 20150013

四、 **Product Registration Certificate No, Product Technical Requirements No:** HNFDA Certified 20182140023

## 五、 Structure and composition, scope of application

The nursing type consists of basic configuration tray, plastic tweezers, medical gauze block, auxiliary configuration paper towel, medical skimmed cotton ball, absorbent pipe and disposable tongue pressing board; the examination type consists of basic configuration tray, plastic tweezers, mouth mirror, probe, iodine volt cotton ball and auxiliary configuration disinfection pliers and alcohol cotton Ball, scarf composition. The product was sterilized by ethylene oxide and provided sterile.

The products are used for clinical patients and patients in medical institutions for oral examination and nursing. The iodine cotton balls and alcohol cotton balls are suitable for disinfection of hands, skin and mucous membrane.

六、 **Use method:** open the bag can be used

## 七、 Contraindications, precautions, warnings and reminders

Product contraindications: None

Damaged packaging prohibited use, expired products prohibited use;

Disposable use, please destroy after use

Avoid high temperature, humidity, direct sunlight

## 八、 Interpretation of graphics, symbols, abbreviations and other contents used in medical device labels



一次性使用



参考使用说明书



包装破损切勿使用



环氧乙烷灭菌



无菌

## 九、 Installation and use instructions or illustrations

This product does not need to be installed, open the bag can be used.

## 十、 Product maintenance and maintenance methods, special storage, transportation conditions, methods

Products should be stored in relative humidity  $\leq 80\%$ , non-corrosive gas and well-ventilated indoor, ordinary transport, in the transportation, storage process should be kept clean and pollution-free, strictly prohibited rain and snow immersion.

## 十一、 Date of production, shelf life or expiration date

Valid for two years, production batch number, production date, expiry date please refer to the small package or qualification label.

**十二、 Accessories list, including accessories, appendix, lossitage replacement cycle and replacement method description**

This product has n o accessories

**十三、 Date and version of preparation or revision of the specification**

This specification was revised in February 2023, version 2023.